



*Delivering more than
just test results*

ALG ORELAP ID #OR100012
361 West 5th Ave
Eugene, OR 97401
TEL: (541) 485-8404 FAX: (541) 484-5995
Website:

September 23, 2016

Doug Pigman
Greater Albany Public Schools
3610 Grand Prairie
Albany, OR 97322
TEL: (541) 967-4513
FAX

RE: NAMS Retest

Order No.: 1609928

Dear Doug Pigman:

Analytical Laboratory Group received 7 sample(s) on 9/20/2016 for the analyses presented in the following report.

A handwritten signature in black ink that reads "Kimberly J. Reeve Morghan".

Kimberly Reeve Morghan
Quality Manager
361 West 5th Ave
Eugene, OR 97401



*Delivering more than
just test results*

ALG ORELAP ID #OR100012

361 West 5th Ave

Eugene, OR 97401

TEL: (541) 485-8404 FAX: (541) 484-5995

Website:

Case Narrative

WO#: **1609928**

Date: **9/23/2016**

CLIENT: Greater Albany Public Schools

Project: NAMS Retest

This report presents the results of the analyses of the sample(s) received on the date above and assigned the listed ALG lab report numbers. Test results relate only to the parameters tested and to the samples as received by the laboratory.

This report shall not be reproduced, except in full, without written consent of Analytical Laboratory Group, Inc.

All analyses were performed according to the Analytical Laboratory Group, Inc. Quality Assurance Program.

All QA/QC requirements were met except as noted below.

Analytical comments are noted with data flags on the reports and/or below.

WO#: 1609928
CLIENT: Greater Albany Public Schools
Project: NAMS Retest
PWS Number:
Sample Source:

Received Date: 9/20/2016 2:40:00 PM
Sampler Name: Jessica Dilbone
Matrix: Drinking Water
Sample Type:

Lab ID: 1609928-001 **Client Sample ID:** Rm 19 SDF **Collection Date:** 9/15/2016 6:13:00 AM

Analyses	Method	Result	MCL	RL	Qual	Units	Date Analyzed	Analys
Lead	SM 3113 B	0.00215	0.0200	0.00200		mg/L	9/22/2016 7:24:00 AM	PG

Lab ID: 1609928-002 **Client Sample ID:** Rm 21 Faucet 1 **Collection Date:** 9/15/2016 6:16:00 AM

Analyses	Method	Result	MCL	RL	Qual	Units	Date Analyzed	Analys
Lead	SM 3113 B	0.00505	0.0200	0.00200		mg/L	9/22/2016 7:24:00 AM	PG

Lab ID: 1609928-003 **Client Sample ID:** Bookroom Faucet **Collection Date:** 9/15/2016 6:18:00 AM

Analyses	Method	Result	MCL	RL	Qual	Units	Date Analyzed	Analys
Lead	SM 3113 B	0.0227	0.0200	0.00400	*	mg/L	9/22/2016 7:24:00 AM	PG

Lab ID: 1609928-004 **Client Sample ID:** Storeroom Faucet **Collection Date:** 9/15/2016 6:20:00 AM

Analyses	Method	Result	MCL	RL	Qual	Units	Date Analyzed	Analys
Lead	SM 3113 B	0.0237	0.0200	0.00400	*	mg/L	9/22/2016 7:24:00 AM	PG

Lab ID: 1609928-005 **Client Sample ID:** Home Ec Faucet 1 **Collection Date:** 9/15/2016 6:08:00 AM

Analyses	Method	Result	MCL	RL	Qual	Units	Date Analyzed	Analys
Lead	SM 3113 B	0.00604	0.0200	0.00200		mg/L	9/22/2016 7:24:00 AM	PG

Lab ID: 1609928-006 **Client Sample ID:** Home Ec Faucet 6 **Collection Date:** 9/15/2016 6:09:00 AM

Analyses	Method	Result	MCL	RL	Qual	Units	Date Analyzed	Analys
Lead	SM 3113 B	0.00614	0.0200	0.00200		mg/L	9/22/2016 7:24:00 AM	PG

Lab ID: 1609928-007 **Client Sample ID:** Home Ec Faucet 7 **Collection Date:** 9/15/2016 6:10:00 AM

Analyses	Method	Result	MCL	RL	Qual	Units	Date Analyzed	Analys
Lead	SM 3113 B	0.00848	0.0200	0.00200		mg/L	9/22/2016 7:24:00 AM	PG

Qualifiers:

*	Value exceeds Maximum Contaminant Level (MCL)	A	Accredited by ORELAP
C	Value is below Minimum Compound Limit.	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	LOD	Limit of Detection
MCL	Maximum Contaminant Level	NAR	See note in Case Narrative
ND	Not Detected at the Reporting Limit	PL	Permit Limit



Delivering more than just test results

ALG ORELAP ID #OR100012

361 West 5th Ave

Eugene, OR 97401

TEL: (541) 485-8404 FAX: (541) 484-5995

Website:

Accreditation Program Analytes Report

WO#: 1609928

23-Sep-16

Client: Greater Albany Public Schools

Project: NAMS Retest

Program Name	Sample ID	ClientSampleID	Matrix	Test Name	Analyte	Status
ORELAP	1609928-001A	Rm 19 SDF	Drinking Water	AA Metals by SM 3113 Schools 250mL	Lead	A
	1609928-002A	Rm 21 Faucet 1			Lead	A
	1609928-003A	Bookroom Faucet			Lead	A
	1609928-004A	Storeroom Faucet			Lead	A
	1609928-005A	Home Ec Faucet 1			Lead	A
	1609928-006A	Home Ec Faucet 6			Lead	A
	1609928-007A	Home Ec Faucet 7			Lead	A

ORELAP A Accredited

ACCRED

Original #1609928# v1

Page 4 of 6



Delivering more than
just test results

ALG ORELAP ID #OR100012

361 West 5th Ave

Eugene, OR 97401

TEL: (541) 485-8404 FAX: (541) 484-5995

Website:

Definition Base

WO#: 1609928

Date: 9/23/2016

Definitions:

% REC: Percent Recovery; a measure of accuracy expressed as a percentage of a measured (recovered) concentration compared to the known concentration added to the sample.

% RPD: Relative Percent Difference; a measure of precision expressed as a percentage of the difference between two duplicates relative to the average concentration.

DF: Dilution factor; the dilution factor applied to the prepared sample.

DUP: Duplicate; aliquots of a sample taken from the same container under laboratory conditions and processed and analyzed independently, used to calculate Precision (%RPD).

LCS: Laboratory Control Sample; prepared by adding a known mass of target analytes to a specified amount of de-ionized water and prepared with the batch of samples, used to calculate Accuracy (%REC).

LCSD: The duplicate sample of the LCS, used to calculate both Accuracy (%REC) and Precision (%RPD)

MBLK: Method Blank; a sample of similar matrix that is processed simultaneously with and under the same conditions as samples through all steps of the analytical procedure, and in which no target analytes or interferences are present at concentrations that impact the analytical results for sample analyses.

MS: Matrix Spike; prepared by adding a known mass of target analytes to a specified amount of matrix sample for which an independent estimate of target analyte concentration is available, used to calculate Accuracy (%REC)

MSD: The duplicate sample of the MS, used to calculate both Accuracy (%REC) and Precision (%RPD)

ND: Not Detected. The analyte level is below the lowest point the laboratory can test for.

PL: Permit limit; only applicable to wastewater reports.

PQL: Practical Quantitation Level or Reporting Limit; the limit to which data is compared for reporting.

Qual: Qualifier that applies to the analyte reported

Definitions:

Result: Analyte concentration reported

RL: Reporting Limit/Limit of Quantitation; the limit to which data is compared for reporting. Analyte concentrations below the reporting limit are reported as ND or with a “J” qualifier.

Units: The units in which the analyte concentration is reported.

Qualifiers:

- * Value exceeds Maximum Contaminant Level (MCL)
- A Accredited by ORELAP
- C Value is below Minimum Compound Limit.
- E Value above quantitation range
- H Holding times for preparation or analysis exceeded
- LOD Limit of Detection
- MCL Maximum Contaminant Level
- NAR See note in Case Narrative
- ND Not Detected at the Reporting Limit
- PL Permit Limit
- R RPD outside accepted recovery limits
- RL Reporting Detection Limit
- U Samples with CalcVal < MDL
- W Sample container temperature is out of limit as specified at testcode

Analytical Laboratory Group, Inc.

361 WEST FIFTH AVENUE
 EUGENE, OREGON 97401
 800-262-5973/541-485-8404 Fax 541-484-5995
 Email: alglabs@alglabsinc.com



Delivering more
 just test rest

NPUC

CHAIN OF CUSTODY

Attention: Doug Pigman	Client: Greater Albany Public Schools
Phone: 541-967-4513	Address: 3610 Grand Prairie
Email: doug.pigman@albany.k12.or.us	City, State, Zip: Albany, OR 97322
Client Project: NAMS Retest	Sampler: Print Jessica Dilbone Sampler: Signature Jessica Dilbone

Client ID	Sample Matrix & Grab/Comp	Collection		Analysis Requested	Bottles - Lab Use Only				
		Date	Time		Type	#	Pres	T °C	Lab ID
Rm 19 SDF	DW/Grab	9-15-16	6:13a	Lead	P	1			001A
Rm 21 Faucet 1	DW/Grab	9-15	6:16a	Lead	P	1			002A
Bookroom Faucet	DW/Grab	9-15	6:18a	Lead	P	1			003A
Storeroom Faucet	DW/Grab	9-15	6:20a	Lead	P	1			004A
Home Ec Faucet 1	DW/Grab	9-15	6:08a	Lead	P	1			005A
Home Ec Faucet 6	DW/Grab	9-15	6:09a	Lead	P	1			006A
Home Ec Faucet 7	DW/Grab	9-15	6:10a	Lead	P	1			007A
	DW/Grab			Lead	P	1			
	DW/Grab			Lead	P	1			
	DW/Grab			Lead	P	1			

Notes:	Preservation Check				
	Lab ID	Date/Time	Pre-Preserved	pH	Tech

Turn Around Time Requested (Rush incurs a Surcharge): <input type="checkbox"/> NORMAL _____ <input checked="" type="checkbox"/> RUSH _____	Shipped Via: ALG Courier	Refrigerated NA
--	------------------------------------	---------------------------

Relinquished by: Jessica Dilbone	Date 9-20-16	Time 1:31	Received by: [Signature]	Date 9/20/16	Time 1331
Relinquished by: [Signature]	Date 9/20/16	Time 1440	Received by: [Signature]	Date 9/20/16	Time 1440
Relinquished by: [Signature]	Date 9/20/16	Time 1440	Received by Laboratory: [Signature]	Date 9/20/16	Time 1440